EM-06

The Quantitative Diffusion-Tensor Anisotropy of Human Brain Using Fast STEAM DTI

박현정¹ • 황문정² • 김용선¹ • 이상권¹ • 장용민^{1,2} ¹경북대 의과대학 진단방사선과학교실 ²경북대 대학원 의공학과

- 목적: To obtain quantitative diffusion-tensor anisotropy information of human cerebral structu using turbo STEAM diffusion-tensor imaging.
- 대상 및 방법: Quantitative diffusion anisotropy MR images were obtained in 7 healthy adults using turbo STEAM sequence and a combination of tetrahedral and orthogonal diffusi gradients. Both relative anisotropy(RA) and fractional anisotropy(FA) values were measured various brain regions. The anisotropy index was then compared with the reported value resulting from EPI-based diffusion tensor imaging.
- 결과: Anisotropy maps depicted white matter features not typically seen on conventional M images. Anisotropy differences, which were statistically significant, occurred across differe WM regions. It is also shown that FA index is higher than RA index.
- 결론: The fact that statistically significant differences were found between anisotropy values different classes of WM suggests that diffusion tensor measurement has a sensitivity microscopic WM structure and may have clinical implication to identify WM disease.