

Original Article

Clinical Experience with 3.0 T MR for Cardiac Imaging in Patients: Comparison to 1.5 T using Individually Optimized Imaging Protocols

Jeong Min Ko¹, Jung Im Jung², Bae Young Lee³

¹Department of Radiology, St. Vincent's Hospital, College of Medicine, The Catholic University of Korea, Suwon, Korea

²Department of Radiology, Seoul St. Mary's Hospital, College of Medicine, The Catholic University of Korea, Seoul, Korea

³Department of Radiology, St. Paul's Hospital, College of Medicine, The Catholic University of Korea, Seoul, Korea

Purpose : To report our clinical experience with cardiac 3.0 T MRI in patients compared with 1.5 T using individually optimized imaging protocols.

Materials and Methods: We retrospectively reviewed 30 consecutive patients and 20 consecutive patients who underwent 1.5 T and 3 T cardiac MRI within 10 months. A comparison study was performed by measuring the signal-to-noise ratio (SNR), the contrast-to-noise ratio (CNR) and the image quality (by grading each sequence on a 5-point scale, regarding the presence of artifacts).

Results: In morphologic and viability studies, the use of 3.0 T provided increase of the baseline SNRs and CNRs, respectively (T1: SNR 29%, $p < 0.001$, CNR 37%, $p < 0.001$; T2-SPAIR: SNR 13%, $p = 0.068$, CNR 18%, $p = 0.059$; viability imaging: SNR 45%, $p = 0.017$, CNR 37%, $p = 0.135$) without significant impairment of the image quality (T1: 3.8 ± 0.9 vs. 3.9 ± 0.7 , $p = 0.438$; T2-SPAIR: 3.8 ± 0.9 vs. 3.9 ± 0.5 , $p = 0.744$; viability imaging: 4.5 ± 0.8 vs. 4.7 ± 0.6 , $p = 0.254$). Although the image qualities of 3.0 T functional cine images were slightly lower than those of 1.5 T images (3.6 ± 0.7 vs. 4.2 ± 0.6 , $p < 0.001$), the mean SNR and CNR at 3.0 T were significantly improved (SNR 143% increase, CNR 108% increase, $p < 0.001$). With our imaging protocol for 3.0 T perfusion imaging, there was an insignificant decrease in the SNR (11% decrease, $p = 0.172$) and CNR (7% decrease, $p = 0.638$). However, the overall image quality was significantly improved (4.6 ± 0.5 vs. 4.0 ± 0.8 , $p = 0.006$).

Conclusion: With our experience, 3.0 T MRI was shown to be feasible for the routine assessment of cardiac imaging.

Index words : 3.0 Tesla · Cardiac MR · SNR · CNR · Image quality

INTRODUCTION

Cardiovascular magnetic resonance imaging (MRI)

• Received; July 22, 2012 • Revised; April 1, 2013

• Accepted; April 1, 2013

Corresponding author : Jung Im Jung, M.D.

Department of Radiology, Seoul St. Mary's Hospital, College of Medicine, The Catholic University of Korea, 222 Banpo-daero, Seocho-gu, Seoul 137-701, Korea.

Tel. 82-2-2258-1435, Fax. 82-2-599-6771

E-mail : jijung@catholic.ac.kr

has gained widespread acceptance for the assessment of cardiovascular disease (1–4). Recent developments in 3.0 T MRI technology have led to an increased interest in high-field cardiac imaging. Because the signal-to-noise (SNR) scales approximately linearly with the magnetic field strength, B_0 , an improvement in the SNR can be expected at 3.0 T in comparison to 1.5 T. Previous studies on high-field strength cardiac MR showed a significant increase in the SNR but also reported image quality problems that were associated with B_1 field inhomogeneities, specific absorption rate (SAR) constraints and a pronounced sensitivity to

susceptibility artifacts (5, 6). However, improvements in MRI hardware and the development of tailored imaging techniques have been cornerstones for rapid progress in cardiac MRI.

Recently, preliminary studies on the feasibility of high magnetic field strengths for cardiac MRI compared to 1.5 T in healthy volunteers using nearly identical imaging protocols (i.e., effect of high magnetic field strength on cardiac MRI) have been reported (7-9). However, there has been rarely reported that direct comparison of 3.0 T cardiac MRI to 1.5 T cardiac MRI in clinical practice without controlling the MR parameters, which is more applicable in clinical settings, because each 1.5 T and 3.0 T protocol has its own idealized parameters set up to maximized good image quality. The aim of this study is to report our clinical experience with cardiac 3.0 T MRI compared to 1.5 T MRI in patients using standard clinically acceptable imaging protocols.

MATERIALS AND METHODS

Study populations

Our retrospective study was approved by the local

institutional review board and used a waiver of informed consent. A total of 50 patients (35 males and 15 females; mean age of 51 years \pm 16 years; age range from 18-80 years) who were referred for MRI assessment of clinically suspected heart disease were enrolled in this study. We retrospectively reviewed 30 consecutive patients (myocardial infarction [n = 14], normal [n = 8], non-ischemic cardiomyopathy [n = 5], valvular heart disease [n = 1], myocarditis [n = 1], congenital heart disease [n = 1]) who underwent 1.5 T cardiac MRI and 20 consecutive patients (myocardial infarction [n = 7], normal [n = 6], non-ischemic cardiomyopathy [n = 4], cardiac neoplasm [n = 2], valvular heart disease [n = 1]) who underwent 3.0 T cardiac MRI between April 2009 and January 2010 (Table 1). The diagnoses were based upon compatible clinical, laboratory, and echocardiographic features. A case of myocarditis and two cases of cardiac neoplasms were pathologically confirmed. The patients were randomly distributed to first available machine at the time of the MRI order. In all patients, impaired breath-holding capacity or arrhythmia that effects on acquisition of MRI was not detected, and the MRI were successfully completed.

Table 1. Characteristics of study population

Characteristic	3.0 T	1.5 T
No. of patients	20	30
Age (range)	47.3 \pm 16.6 (18-73)	53.0 \pm 15.8 (23-80)
Sex (M/F)	14/6	21/9
Diagnosed cardiac disease		
Myocardial infarction	7	14
Nonischemic cardiomyopathy	4 Dilated cardiomyopathy (n = 4)	5 Dilated cardiomyopathy (n = 3) Hypertrophic cardiomyopathy (n = 2)
Valvular heart disease	1 Pulmonary stenosis (n = 1)	1 Aortic regurgitation (n = 1)
Myocarditis	0	1
Cardiac neoplasm	2 Angiosarcoma (n = 1) Myxoma (n = 1)	0
Congenital heart disease	0	1 Atrial septal defect (n = 1)
Normal	6	8

MR imaging

All cardiac MRI examinations were performed using a 1.5 T MR scanner (Intera Achieva; Philips Medical Systems, Best, The Netherlands) or a 3.0 T MR scanner (Magnetom Verio; Siemens Medical solution, Erlangen, Germany) with cardiac-phased array coils and echocardiography gating.

For the morphology study, axial T1-weighted fast spin-echo (1.5 T: TR = 705.9 ms, TE = 10 ms, slice thickness = 8 mm, interslice gap = 2 mm, field of view = 320 × 320 mm², matrix = 320 × 293; 3.0 T: TR = 812.5 ms, TE = 8.6 ms, slice thickness = 8 mm, interslice gap = 2 mm, field of view = 300 × 300 mm², matrix = 256 × 179) and T2-weighted Spectral Attenuated (Spectrally Adiabatic) Inversion Recovery (T2-SPAIR, 1.5 T: TR = 1411.8 ms, TE = 100 ms, slice thickness = 8 mm, interslice gap = 2 mm, field of view = 320 × 320 mm², matrix = 236 × 151; 3.0 T: TR = 769 ms, TE = 52 ms, slice thickness = 8 mm, interslice gap = 2 mm, field of view = 300 × 300 mm², matrix = 256 × 205) sequences were used. Data acquisition for each slice was completed in separate breath-holds.

For the functional cine imaging, a balanced steady-state free precession (SSFP) imaging sequence was performed using a short axis 2-chamber view, a long axis 2-chamber view and a 4-chamber view (1.5 T: TR = 3.3 ms, TE = 1.6 ms, flip angle = 60°, slice thickness = 8 mm, interslice gap = 2, field of view = 320 × 320 mm², matrix = 192 × 192; 3.0 T: TR = 3.26 ms, TE = 1.4 ms, flip angle = 38°, slice thickness = 8 mm, interslice gap = 2, field of view = 298 × 300 mm², matrix = 256 × 178).

After the injection of 0.1 mmol/kg of gadolinium-DTPA (Magnevist, Schering, Berlin, Germany) at a rate of 4 ml/s and a 20 ml normal saline flush using a power injector (Spectris, Medrad, Indiaola, PA), the first pass perfusion images were obtained using a balanced steady-state free precession (SSFP) imaging sequence at 1.5 T and a saturation-recovery (SR) gradient recalled echo sequence (turbo FLASH) at 3 T. A dynamic series (40–60 frames) of three, short axis slices (apical, mid-ventricular and basal) per R-R interval was continuously acquired (1.5 T: TR = 2.7 ms, TE = 1.3 ms, flip angle = 50°, slice thickness = 8 mm, field of view = 320 × 320 mm², matrix = 128 × 102; 3.0 T: TR = 1.94 ms, TE = 1.1 ms, flip angle = 10°, slice thickness = 8 mm, field of view = 300 ×

300 mm², matrix = 192 × 154).

Followed by first pass perfusion image, the repeated bolus injection of 0.1 mmol/kg of gadolinium-DTPA was done. Delayed enhancement images for viability study were acquired 10 minutes after contrast injection using an inversion-recovery spoiled gradient-echo sequence (1.5 T: TR = 5.0 ms, TE = 1.5 ms, flip angle = 15°, slice thickness = 8 mm, interslice gap = 2 mm, field of view = 320 × 320 mm², matrix = 256 × 256; 3.0 T: TR = 5.34 ms, TE = 2 ms, flip angle = 20°, slice thickness = 8 mm, interslice gap = 2 mm field of view = 300 × 300 mm², matrix = 256 × 205) through the whole left ventricle on the short-axis view.

Data analysis

Two radiologists independently and blindly measured the SNR and CNR of each sequence. For the T1-weighted, T2-SPAIR, functional cine and first pass perfusion images, the signal intensity of the myocardium and blood were defined as the mean signal from circular 20–30 mm² ROIs of the normal myocardium at the middle of the left ventricle. The signal intensity of the blood was defined as the mean signal intensity of a circular 50–60 mm² ROI of the left ventricular cavity. For the first-pass perfusion imaging, the SNR was determined for the images that showed the maximum intensity during the contrast agent passage through the left ventricular myocardium. The SNR and CNR were calculated as the SNR = mean signal intensity of myocardium/noise and the CNR = |mean signal intensity of blood - mean signal intensity of myocardium|/noise.

For the viability imaging, the SNR was obtained from ROIs positioned in the left ventricular myocardial areas that showed a delay in the contrast agent accumulation. The SNR and CNR were computed as SNR = mean signal intensity of the myocardium with delayed enhancement/noise and the CNR = |mean signal intensity of the myocardium with delayed enhancement - mean signal intensity of the normal myocardium|/noise.

Although the conventional measure of noise is the standard deviation of the signal intensity in regions outside of the body, parallel imaging reconstruction algorithms induce a very inhomogeneous noise amplification that result in a generally non-uniform background noise distribution (10). To avoid misrepre-

senting the actual SNR and CNR calculations, we used the statistics from the air space in the lungs as a surrogate measure of noise (11).

The image quality was graded by two radiologists in consensus using a 5-point scale (score 5 = very good/no artifacts; 4 = good/only a few artifacts; 3 = fair/artifacts are present but did not influence the image quality; 2 = poor/artifacts are present and influence the image quality; 1 = very poor/not assessable as a result of artifacts).

Statistical analysis

For both field strengths, the mean and the mean standard deviations were calculated for the data derived from the SNRs, CNRs and image scores. The independent t-test was used to assess the statistical significances. Interobserver variability for measurements of the SNR and CNR of each sequence was analyzed by calculating the intraclass correlation coefficient (ICC) with the two-way random effects model. The ICC results for interobserver agreement were interpreted using the ICC interpretation guideline (12). According to this guideline, an ICC value lower than 0.4 suggests that the observers are in poor agreement. An ICC value between 0.4 and 0.75 suggests that the level of agreement is fair to good, while an ICC value greater than 0.75 suggests excellent agreement. In all tests, p values less than 0.05 were considered statistically significant.

RESULTS

The results from the quantitative analysis are provided in Table 2. In morphologic and viability studies, the use of 3.0 T provided significant and insignificant improvements in the baseline SNRs and CNRs, respectively (T1: SNR 29% increase, $p < 0.001$, CNR 37% increase, $p < 0.001$; T2-SPAIR: SNR 13% increase, $p = 0.068$, CNR 18% increase, $p = 0.059$, viability imaging: SNR 45% increase, $p = 0.017$, CNR 37% increase, $p = 0.135$) without significant impairment in the image qualities (T1: 3.8 ± 0.9 with 3.0 T versus 3.9 ± 0.7 with 1.5 T, $p = 0.438$; T2-SPAIR: 3.8 ± 0.9 versus 3.9 ± 0.5 , $p = 0.744$; viability imaging: 4.5 ± 0.8 versus 4.7 ± 0.6 , $p = 0.254$).

Although the image qualities from the 3.0 T functional cine imaging were slightly lower than those of the 1.5 T (3.6 ± 0.7 on 3.0 T versus 4.2 ± 0.6 on 1.5 T, $p < 0.001$), the mean SNR and CNR with the 3.0 T were significantly improved (SNR 143% increase, $p < 0.001$, CNR 108% increase, $p < 0.001$).

With our imaging protocol for 3.0 T perfusion imaging, there were insignificant decreases in the SNR and CNR (SNR 11% decrease, $p = 0.172$; CNR 7% decrease, $p = 0.638$). However, the overall image quality was significantly improved (4.6 ± 0.5 versus 4.0 ± 0.8 , $p = 0.006$) due to the decrease of dark rim artifacts with 3.0 T MRI.

Table 2. SNR, CNR, and Image Quality at 3.0 T and 1.5 T in Cardiac MR

			3.0 T	1.5 T	p value
Morphology imaging	T1WI	SNR	49.8 ± 12.0	38.5 ± 6.9	< 0.001
		CNR	41.6 ± 11.5	30.3 ± 5.3	< 0.001
		Image quality	3.8 ± 0.9	3.9 ± 0.7	0.438
	T2-SPAIR	SNR	42.2 ± 12.4	37.2 ± 6.1	0.068
		CNR	38.0 ± 11.5	32.0 ± 6.1	0.059
		Image quality	3.8 ± 0.9	3.9 ± 0.5	0.744
Functional CINE imaging	SNR	70.4 ± 27.3	29.0 ± 11.8	< 0.001	
	CNR	144.0 ± 73.4	69.2 ± 27.8	< 0.001	
	Image quality	3.6 ± 0.7	4.2 ± 0.6	< 0.001	
First-pass perfusion imaging	SNR	27.9 ± 6.6	31.3 ± 9.7	0.172	
	CNR	35.5 ± 15.9	38.0 ± 19.7	0.638	
	Image quality	4.6 ± 0.5	4.0 ± 0.8	0.006	
Viability imaging	SNR	52.9 ± 16.7	36.6 ± 11.8	0.017	
	CNR	38.4 ± 19.4	28.0 ± 11.5	0.135	
	Image quality	4.5 ± 0.8	4.7 ± 0.6	0.254	

Table 3. Interobserver Agreement (Intraclass Correlation Coefficient, ICC) for Measurements Made by Two Radiologists

			3.0 T		1.5 T	
			ICC	p value	ICC	p value
Morphology imaging	T1WI	SNR	0.695	0.014	0.752	0.028
		CNR	0.478	0.082	0.508	0.141
	T2-SPAIR	SNR	0.982	< 0.001	0.830	0.011
		CNR	0.983	< 0.001	0.796	0.017
Functional CINE imaging		SNR	0.916	0.001	0.961	0.002
		CNR	0.918	0.002	0.869	0.015
First-pass perfusion imaging		SNR	0.695	0.031	0.457	0.164
		CNR	0.863	0.008	0.686	0.042
Viability imaging		SNR	0.855	0.021	0.761	0.054
		CNR	0.920	0.007	0.929	0.005

Table 3 shows the degree of agreement between the measurements of two radiologists using the ICC method. Interobserver agreement was excellent for measurements of SNRs and CNRs for T2-SPIR, functional cine, and viability imagings of 3.0 T and 1.5 T, CNR of 3.0 T perfusion imaging, and SNR of 1.5 T T1WI. Fair to good agreement was observed for SNRs and CNRs of 3.0 T T1WI and 1.5 T perfusion imaging, SNR of 3.0 T perfusion imaging, and CNR of 1.5 T T1WI.

DISCUSSION

Our results on morphologic, viability and functional cine studies are similar to previous preliminary studies that used nearly identical imaging protocols for 1.5 T and 3.0 T MRI to evaluate the effect of high magnetic fields (7–9). Our results show that, for morphologic, functional cine, and viability images, the use of different parameters for clinically optimized imaging protocols for 3.0 T compared to 1.5 T have little influence on the SNR, CNR, and image quality. Therefore, as with preliminary studies on the feasibility of high magnetic strength cardiac MR, our clinical experience suggests that cardiac MR at 3.0 T is feasible for the assessment of cardiac morphology, viability, and function.

Theoretically, imaging at 3.0 T leads to a 2-fold increase in the SNR compared to 1.5 T as the signal increases with B_0 (3). The higher SNR at 3.0 T can

also be used to increase the spatial resolution and/or reduce the imaging time by means of parallel imaging, which means under-sampling the k-space in conjunction with multi-element coils for spatially resolved signal detection (10, 13). Morphologic studies using fast spin-echo based imaging sequences benefit from the synergy between high magnetic field strengths and parallel imaging.

In viability studies, the delayed enhancement techniques that were clinically established at 1.5 T offer a relatively low SNR and very limited spatial coverage with a maximum of 3 slices per breath-hold. Parallel imaging at 3.0 T can overcome these difficulties by allowing whole-heart coverage in a single breath hold, leading to a uniform suppression of healthy myocardium in all of the imaged sections and increased patient comfort (7). With regard to the contrast between the normal and infarcted myocardium at 3.0 T, there is an increase in the CNR in contrast-enhanced T1-weighted sequences between the infarcted and normal myocardium because of the higher sensitivity to the gadolinium-based contrast agents used at 3.0 T (14). The prolongation of the T1-relaxation times at 3.0 T is expected to improve the tissue enhancing properties of the T1-shortening contrast agents, such as gadolinium-DTPA.

Strong static field inhomogeneities and short T2* cause more artifacts with SSFP cine imaging at 3 T than at lower field strengths (15). Banding artifacts are one of the most commonly observed artifacts in 3.0 T SSFP imaging; these artifacts can reduce image quality

and hamper diagnosis of cardiac disease (15, 16). Minimization of these limitation could be achieved with localized linear or second-order shimming and a frequency scout sequence (15, 17). Although the image quality of functional cine images at 3.0 T in this study was slightly inferior due to artifacts, the diagnostic performance was not impaired (the image quality at 3.0 T: 3.6 ± 0.7).

In our study, first-pass perfusion imaging at 3.0 T showed an insignificant decreased in the SNR and CNR and significant improved image quality. The presence of dark rim artifacts is one of the major weaknesses of myocardial MR perfusion imaging. Typically, these artifacts are most pronounced when the gadolinium contrast bolus appears in the left ventricular cavity preceding tissue uptake of contrast agent and are transient in nature (18). The potential causes of dark rim artifacts are currently postulated to be the result of factors such as the gadolinium bolus (18), the motion of cardiac cycle (19), and limited spatial resolution that results in Gibbs rings (20). Strach et al. reported that high-resolution myocardial perfusion imaging at 3.0 T with a 60% increase in spatial resolution compared to the standard clinical perfusion protocol at 1.5 T is feasible and results in a significant improvement in the overall image quality and a significant reduction in dark rim artifacts (21). In another comparison study of perfusion imaging at 1.5 T and 3.0 T using identical voxel size (22), the result suggested that the diagnostic performance of 3.0 T perfusion imaging is significantly greater than that of 1.5 T in identifying both single-vessel disease and multivessel disease. In our study, small fields of view (FOV) and large matrices were used for a first-pass perfusion study at 3.0 T (FOV = $300 \times 300 \text{ mm}^2$, matrix = 192×154 , pixel size = 3.04 mm^2 at 3.0 T versus FOV = $320 \times 320 \text{ mm}^2$, matrix = 128×102 , pixel size = 7.84 mm^2 at 1.5 T). We believe that high spatial resolution (i.e., smaller pixel sizes) at 3.0 T contribute to the improvement in image quality by decreasing the occurrence of dark rim artifacts (7 of 20 patients (35%) at 3.0 T versus 18 of 30 patients (60%) at 1.5 T). We used a turbo FLASH sequence for first-pass perfusion study at 3.0 T whereas balanced SSFP sequence was used at 1.5T. SSFP sequence have been proven to provide high SNR and CNR and to shorten acquisition time compared to gradient-echo

sequences (23). We thought that use of turbo FLASH sequence at 3.0 T probably caused decrease of SNR and CNR as compared with those of 1.5 T in spite of advantage of high magnetic field.

Our study had some limitations. The relatively small size of patient group with diverse diseases is a major limitation. Use of different sequences for first-pass perfusion imaging could also affect the result of this study.

CONCLUSION

In our clinical experience, 3.0 T MRI was feasible for the routine assessment of cardiac images, although the results in functional studies were slightly inferior. Moreover, in perfusion imaging, 3.0 T provided improved diagnostic value by improving the image quality.

References

1. Dianas PG, Stuber M, Botnar RM, et al. Coronary MR angiography clinical applications and potential for imaging coronary artery disease. *Magn Reson Imaging Clin N Am* 2003;11:81-99
2. Edelman RR. Contrast-enhanced MR Imaging of the Heart: Overview of the Literature. *Radiology* 2004;232:653-668
3. Lima JA, Desai MY. Cardiovascular magnetic resonance imaging: current and emerging applications. *J Am Coll Cardiol* 2004;44:1164-1171
4. Pennell DJ, Sechtem UP, Higgins CB, et al. Clinical indications for cardiovascular magnetic resonance (CMR): consensus Panel report. *Eur Heart J* 2004;25:1940-1965
5. Hinton DP, Wald LL, Pitts J, Schmitt F. Comparison of cardiac MRI on 1.5 and 3.0 Tesla clinical whole body systems. *Invest Radiol* 2003;38:436-442
6. Gutberlet M, Schwinge K, Freyhardt P, et al. Influence of high magnetic field strengths and parallel acquisition strategies on image quality in cardiac 2D CINE magnetic resonance imaging: comparison of 1.5 T vs. 3.0 T. *Eur Radiol* 2005;15:1586-1597
7. Gutberlet M, Noeske R, Schwinge K, Freyhardt P, Felix R, Niendorf T. Comprehensive cardiac magnetic resonance imaging at 3.0 Tesla: feasibility and implications for clinical applications. *Invest Radiol* 2006;41:154-167
8. Michaely HJ, Nael K, Schoenberg SO, et al. Analysis of cardiac function--comparison between 1.5 Tesla and 3.0 Tesla cardiac cine magnetic resonance imaging: preliminary experience. *Invest Radiol* 2006;41:133-140
9. Ligabue G, Fiochi F, Ferraresi S, et al. 3-Tesla MRI for the evaluation of myocardial viability: a comparative study with 1.5-Tesla MRI. *Radiol Med* 2008;113:347
10. Pruessmann KP, Weiger M, Scheidegger MB, Boesiger P. SENSE: sensitivity encoding for fast MRI. *Magn Reson Med* 1999;42:952-962

11. Mascarenhas NB, Muthupillai R, Cheong B, Pereyra M, Flamm SD. Fast 3D cine steady-state free precession imaging with sensitivity encoding for assessment of left ventricular function in a single breath-hold. *AJR Am J Roentgenol* 2006;187:1235-1239
12. Rosner B. *Fundamentals of biostatistics*. 6th ed. Belmont, CA: Duxbury Press, 2006:613-617
13. Sodickson DK, Manning WJ. Simultaneous acquisition of spatial harmonics (SMASH): fast imaging with radiofrequency coil arrays. *Magn Reson Med* 1997;38:591-603
14. Klumpp B, Fenchel M, Hoevelborn T, et al. Assessment of myocardial viability using delayed enhancement magnetic resonance imaging at 3.0 Tesla. *Invest Radiol* 2006;41:661-667
15. Noeske R, Seifert F, H. RK, Rinneberg H. Human cardiac imaging at 3 T using phased array coils. *Magn Reson Med* 2000;44:978-982
16. Reeder SB, Faranesh AZ, Boxerman JL, McVeigh ER. In vivo measurement of T_2^* and field inhomogeneity maps in the human heart at 1.5 T. *Magn Reson Med* 1998;39:988-998
17. Schär M, Kozerke S, Fischer SE, Boesiger P. Cardiac SSFP imaging at 3 Tesla. *Magn Reson Med* 2004;51:799-806
18. Barkhausen J, Hunold P, Jochims M, Debatin JF. Imaging of myocardial perfusion with magnetic resonance. *J Magn Reson Imaging* 2004;19:750-757
19. Storey P, Chen Q, Li W, Edelman RR, Prasad PV. Band artifacts due to bulk motion. *Magn Reson Med* 2002;48:1028-1036
20. Di Bella EV, Parker DL, Sinusas AJ. On the dark rim artifact in dynamic contrast-enhanced MRI myocardial perfusion studies. *Magn Reson Med* 2005;54:1295-1299
21. Strach K, Meyer C, Thomas D, et al. High-resolution myocardial perfusion imaging at 3 T: comparison to 1.5 T in healthy volunteers. *Eur Radiol* 2007;17:1829-1835
22. Cheng ASH, Pegg TJ, Karamitsos TD, et al. Cardiovascular magnetic resonance perfusion imaging at 3-tesla for the detection of coronary artery disease: a comparison with 1.5-tesla. *J Am Coll Cardiol* 2007;49:2440-2449
23. Plein S, Bloomer TN, Ridgway JP, Jones TR, Bainbridge GJ, Sivananthan MU. Steady state free precession magnetic resonance imaging of the heart: comparison with segmented k-space gradient-echo imaging. *J Magn Reson Imaging* 2001;14:230-236

장비 별 최적화된 영상 프로토콜을 이용한 환자에서의 3.0 T 심장 자기공명영상의 임상경험: 1.5 T 자기공명영상과의 비교

¹가톨릭대학교 의과대학 성빈센트병원 영상의학과

²가톨릭대학교 의과대학 서울성모병원 영상의학과

³가톨릭대학교 의과대학 성바오로병원 영상의학과

고정민¹ · 정정임² · 이배영³

목적: 본 연구의 목적은 환자에서 임상적으로 사용되는 영상 프로토콜을 이용하여 시행된 3.0 T 심장자기공명영상을 1.5 T 와 비교하여 그 유용성을 알아보고자 하는데 있다.

대상과 방법: 10개월간 30명의 환자에서 얻은 1.5 T 자기공명영상과 20명의 3.0 T 영상을 후향적으로 비교하였다. 각각의 영상에 대하여 신호 대 잡음비 (signal-to-noise ratio: SNR), 대조도 대 잡음비 (contrast-to-noise ratio: CNR), 영상 화질 (artifact의 정도에 따라서 5단계로 분류)을 평가하고 비교하였다.

결과: T1심장 형태 영상 및 심근 생존능 평가 영상에서는 3.0 T 자기공명영상에서 영상화질 (T1: 3.8 ± 0.9 vs. 3.9 ± 0.7 , $p=0.438$; T2-SPAIR: 3.8 ± 0.9 vs. 3.9 ± 0.5 , $p=0.744$; 지연기 조영 증강 영상: 4.5 ± 0.8 vs. 4.7 ± 0.6 , $p=0.254$)의 유의한 저하 없이 SNR과 CNR의 향상을 보였다 (T1: SNR 29%, $p < 0.001$, CNR 37%, $p < 0.001$; T2-SPAIR: SNR 13%, $p=0.068$, CNR 18%, $p=0.059$; 지연기 조영 증강 영상: SNR 45%, $p=0.017$, CNR 37%, $p=0.135$). 심장Cine 영상에서 3.0 T 심장영상이 1.5 T 영상과 비교하여 영상화질 (3.6 ± 0.7 vs. 4.2 ± 0.6 , $p < 0.001$)이 다소 떨어졌으나 SNR과 CNR의 유의한 상승을 보였다 (SNR 143% 상승, CNR 108% 상승, $p < 0.001$). 심근관류영상에서는 SNR (11% 감소, $p=0.172$)과 CNR (7% 감소, $p=0.638$) 이 통계적으로 유의하지 않은 정도로 감소되었으나 영상화질(4.6 ± 0.5 vs. 4.0 ± 0.8 , $p=0.006$)은 유의한 향상을 보였다.

결론: 실제 임상영역에서 사용되는 영상 프로토콜로 시행된 3.0 T 심장자기공명영상은 1.5T 영상과 비교하여 충분한 영상의 질을 제공하였다.

통신저자 : 정정임, (131-701) 서울시 서초구 반포대로 222, 가톨릭대학교 의과대학 서울성모병원 영상의학과
Tel. (02) 2258-1435 Fax. (02) 599-6771 E-mail: jijung@catholic.ac.kr